

**FLOWMETERS (LIQUID)**

<u>Range</u>	<u>Accuracy</u>	<u>Cal. Device</u>	<u>Notes</u>
-12.2 to 12.2m/s	2 – 5%	Panametrics PT878	Test only.

**AIRFLOW**

<u>Range</u>	<u>Accuracy</u>	<u>Cal. Device</u>	<u>Notes</u>
0 to 60Knots	±1kt	Munro IM159	Rotating cup type.
0.4 to 40m/s	±3% of reading Or ±0.1m/s	Kestrel 1000	25mm Rotating vane type
0.25 to 30m/s	±1% of reading Or ±0.02m/s	TSI Airflow LCA501	100mm Rotating vane type

Last revision date: 18<sup>th</sup> October 2023

Last review date: 18<sup>th</sup> August 2023