

PUMP TESTING

<u>Range</u>	<u>Uncertainty</u>	<u>Cal. Device</u>	<u>Notes</u>
Testing to 60,000psi	0.5%	Test Cabinet	Testing of adaptors, valves, hose assemblies etc

GAS BOOSTER TESTING

<u>Range</u>	<u>Uncertainty</u>	<u>Cal. Device</u>	<u>Notes</u>
Testing to 25,000psi	0.2%	Remote test panel/ Test bay (SDAS)	Testing with industrial Nitrogen

Last revision date: 18th October 2023

Last review date: 18th August 2023